

SPECIFICATIONS

Display:	¾" LCD
Range:	Counts down from 20 hrs.
Resolution:	1 min.
Accuracy:	0.01%
Case:	ABS plastic
Power:	One 3-year AAA battery
Accessories Supplied:	3-year AAA battery, magnet/clip, Traceable® certificate, instructions

QC Timer Operation

1. Turn the timer on by inserting the battery.
To insert the battery, slide the battery cover located on the back of the unit away from the unit. Place the supplied battery inside. Replace the battery cover.
2. To set the desired time, press the HR and MIN buttons. Each press advances the time 1 hour or 1 minute, respectively. To rapidly advance the time, press and hold the buttons.
3. Press the START/STOP button once to begin counting down. The blinking colon (:) indicates the time is running.

4. To temporarily stop the countdown time, press the START/STOP button once. To resume countdown, press the START/STOP button again.
5. When the timer reaches zero, an alarm will sound for 1 minute. Press the ALARM STOP button to silence the sounding alarm.
6. When the timer is counting down, clear the display by pressing the START/STOP button to stop the countdown. Then press the CLEAR button. If the timer is not counting down, press the START/STOP button *twice* and then press CLEAR.

Battery Information

Please replace the battery in the case of any operational difficulties or lack of proper functionality for any reason. Slide the battery cover located on the back of the unit away from the unit. Remove the old battery and replace it with a *new* AAA battery. Replace the battery cover. Incorrectly installed batteries may damage electronics.

Replacing the battery with a new fresh battery will solve most difficulties.

WARRANTY, SERVICE, OR RECALIBRATION

For warranty, service, or recalibration, contact:

CONTROL COMPANY

12554 Old Galveston Rd. Suite B230

Webster, Texas 77598 USA

Ph. 281 482-1714 • Fax 281 482-9448

E-mail sales@control3.com • www.traceable.com

Control Company is ISO 9001:2008 Quality-Certified
by DNV and ISO/IEC 17025:2005 accredited as a
Calibration Laboratory by A2LA.

Traceable® and Ultra™ are registered trademarks/trademarks
of © Control Company.

©2016 © Control Company. 92-5026-00 Rev. 1 082916

**TRACEABLE®
QC TIMER
INSTRUCTIONS**